

# Certificate of Conformity

issued by Curtis-Straus LLC, a wholly owned subsidiary of BV CPS

**Company:** Danatronics Corporation  
**Product Tested:** Ultrasonic Precision Thickness Gage (M/N:  
UPG-07 WAVE)  
**Testing Date:** September 22, 2008  
**Report No.:** ER0188-1

*A sample of the product as configured in the accompanying test report has been found to comply with the following standards:*

## EMC Emissions:

- EN 55022:2010/AC:2011 Class A ITE emissions requirements (EU)
- FCC 47 CFR Part 15 Class A emissions requirements (USA)
- VCCI:V-3/2015.04 Class A ITE emissions requirements (Japan)
- EN 61326-1:2013: EMC requirements for Electrical equipment for measurement, control and laboratory use – General Use


## EMC Immunity:

- EN 61326-1:2013 EMC requirements for Electrical equipment for measurement control and laboratory use – General Use for following tests only:
  - EN 61000-4-2: Electrostatic Discharge
  - EN 61000-4-3: Radiated RF Immunity

*Attested to by the hands and seals:*

For Curtis-Straus LLC

For the Manufacturer or Importer

 2/17/2017  
Test Engineer Date

\_\_\_\_\_  
Date

 2/17/2017  
EMC Project Manager Date